



Calibration
Certificate No. 1750.01

Calibration complies with ISO 9001 ISO/IEC 17025 AND ANSI/NCSL Z540-1



Cert. No.: 4000-3710884

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: Fisher Scientific, P.O. Box 1768, Pittsburgh, PA 15230

Instrument Identification:

Vermont Dept. of Health Lab, 195 Colchester Avenue, Attn. Edmond Luce, Burlington, VT 05402 U.S.A. (RMA:962835)

Model Numbers: 15-077-8, FB50266, 245BY S/N: 240118616 Manufacturer: Control Company

Model: 15-077-7 S/N: 240118587

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-179	A45240		
Thermistor Module	A17118	1/24/12	1000291329
Temperature Probe	128	2/01/12	6-BA02G-1-2
Temperature Calibration Bath TC-231	A79341		
Temperature Probe	3039	2/01/12	6-BA02G-1-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Module	A27129	10/22/11	1000287167
Temperature Probe	5202	4/22/12	6-BD201-1-1
Temperature Calibration Bath TC-256	B01375		
Temperature Probe	157	11/23/11	B0A19013

Certificate Information:

Technician: 68

Procedure: CAL-06

Cal Date: 6/20/11

Cal Due: 6/20/12

Test Conditions: 23.5°C 43.0 %RH 1008 mBar

Calibration Data:

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C	-0.001	0.020	Y	-0.001	0.000	Y	-0.051	0.049	0.013	3.8:1
°C	25.000	24.992	Y	25.000	24.999	Y	24.950	25.050	0.014	3.6:1
°C	60.001	59.989	Y	60.001	60.002	Y	59.951	60.051	0.018	2.8:1
°C	99.999	99.965	Y	99.999	100.002	Y	99.949	100.049	0.013	3.8:1

This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min = Nominal(Rounded) - Tolerance; Max = Nominal(Rounded) + Tolerance; Date=MM/DD/YY

Nicol Rodriguez
Nicol Rodriguez, Quality Manager

Wallace Berry
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).